


SIMPLE LINK - uno





- Provides voltage to the electrospray and connects the column and emitter with an accessible Zero Dead Volume fitting
- Reduce complexity and chances of error.
- Tool-free operation
- Comes in two versions: for IDEX (1/16 OD sleeve, 13 nl swept volume) and VICI (1/32 sleeve, 1 nl swept volume) fittings.
- Compatible with Nano Flex

SIMPLE LINK UNO is the ideal solution for former users metal emitters who want to migrate to the Sharp Singularity™ (mechanically sharpened fused silica emitters).

SIMPLE LINK UNO



- SIMPLE LINK UNO is clamped to the Flex pole.
- The ZDV is comfortably assembled outside, and it simply clicks in.
- SIMPLE LINK UNO collects the voltage from the Flex emitter clamp and passes it to the ZDV fitting and the electrospray. No conductive coating required.
- With no extra wires, SIMPLE LINK provides a clean and robust set-up.




SIMPLE LINK comes in two versions:

- For IDEX ZDV U-411 fittings, with 1/16" sleeves in both sides. Inner diameter: 178 μm , swept volume: 13nL.
- For VICI nano volume fittings, with a 1/32" sleeve on the column side and a 360 μm ferrule on the emitter side. Inner diameter 50 μm , estimated swept volume \sim 1nL


SIMPLE LINK UNO

Install SIMPLE LINK:

1) Slide SIMPLE LINK through the ion source pole



2) Tighten SIMPLE LINK's clamp



3) Connect SIMPLE LINK to Voltage




SIMPLE LINK UNO

Mount the nLC column and Emitter:


1) Assemble the Zero Dead Volume fitting.

Make sure that there are no dead volumes between the sleeves, the emitter and the column. For this, all elements must sit properly in their corresponding position. The Sharp Singularity™ emitters have a straightened and polished back end to ensure a good fit.

SIMPLE LINK facilitates this step by allowing you to mount the ZDV fitting outside the ion source.



2) Click the ZDV assembly into the SIMPLE LINK. .




SIMPLE LINK UNO

Ready to go:

1) Click covers

2) Slide into the source




SIMPLE LINK UNO

Why would you want to integrate the Sharp Singularity™?

To improve the quality of your data
with robust and repeatable nano-electrospray ionization


How the Sharp Singularity emitters improve ionization:

- very acute angle,
- well-defined edges,
- constant ID,
- geometric reproducibility
- full traceability and quality control



Find more information in www.fossiliontech.com

Write to info@fossiliontech.com or your local distributor to order your emitters



FOSSILIONTECH

Proud of servicing science



FOSSILIONTECH

www.fossiliontech.com